

CERTIFICATE OF ANALYSIS No.: 2022-10262

CLIENT

KANNABIO HEMP HELLAS, SKOUFA 110
38334 VOLOS, Greece

SAMPLE *

ACTIVE DROPS 15%

Sample condition: SUITABLE
Sample ID: 2242002
Sample type: Viscous liquid
Batch No.: *Work order: 2022-107003
Analysis ID: 2022_244
Method ID: PHL_RPC_12C
Method SOP: MET-LAB-003-02Sample received: 17/10/2022
Start of analysis: 27/10/2022
End of analysis: 28/10/2022
Analyst: Blaž Janežič

* Information provided by the client.

CANNABINOID TRACE
ANALYSIS

	Concentration [% w/w]	Expanded uncertainty [% w/w]	LOQ [% w/w]	Graphic presentation of relative cannabinoid concentration
CBDV - Cannabidiavin	0.0408	0.0094	0.00030	
CBDA - Cannabidiolic acid	< LOQ	n/a	0.00030	
CBGA - Cannabigerolic acid	< LOQ	n/a	0.00030	
CBG - Cannabigerol	< LOQ	n/a	0.03000	
CBD - Cannabidiol	15.93	0.80	0.03000	█
THCV - Tetrahydrocannabivarin	< LOQ	n/a	0.00030	
CBN - Cannabinol	0.00188	0.00041	0.00030	
Δ⁹-THC - Δ-9-Tetrahydrocannabinol	0.00379	0.00083	0.00030	
Δ⁸-THC - Δ-8-Tetrahydrocannabinol	< LOQ	n/a	0.00030	
CBL - Cannabicyclol	0.00062	0.00014	0.00030	
CBC - Cannabichromene	0.0103	0.0023	0.00030	
Δ⁹-THCA - Δ-9-Tetrahydrocannabinolic acid	< LOQ	n/a	0.00030	
CBE - Cannabielsoin	0.0105 #	0.0029	0.00030	
CBNV - Cannabivarin	< LOQ #	n/a	0.00030	
CBCA - Cannabichromenic acid	< LOQ #	n/a	0.00030	
CBT - Cannabicitran	0.0046 #	0.0010	0.00030	

Units and abbreviations: % w/w = weight percent, LOQ = the limit of quantitation, ND = not detected, n/a = not available.

The results given herein apply only to the sample as received. **Expanded Uncertainty** was calculated using coverage factor k = 2, corresponding to a double standard uncertainty and characterizes the interval value in which it is possible to expect the real value with a probability of 95%. This is stated according to the ISO/IEC Guide 98-3.

Total or partial reproduction of this document is not allowed without the permit from PharmaHemp d.o.o. The document does not substitute any other legal document.

Date issued:

28/10/2022

Approved by:

mag. Marko Dragan
Analytical Laboratory Manager

Authorized by:

dr. Boštjan Jančar
Chief Technology Officer

End of Certificate